



財團法人全國認證基金會
Taiwan Accreditation Foundation

Certificate of Accreditation

(Certificate No : L2845-260107)

This is to certify that

AUO Corporation
Taichung Calibration Laboratory

No. 1, JhongKe Rd. Central Taiwan Science Park, Taichung 40763, Taiwan

is accredited in respect of laboratory

Accreditation Criteria : ISO/IEC 17025:2017 ; CNS 17025:2018

Accreditation Number : 2845

Originally Accredited : February 20, 2014

Effective Period : February 20, 2026 to February 19, 2029

Accredited Scope : Calibration Field, see described in the Appendix



Scan to verify

Yi-Ling Chen

Yi-Ling Chen
President, Taiwan Accreditation Foundation
January 07, 2026

Accreditation Number : 2845

Laboratory Head : CHIU, Yao-Ching

Temperature/Humidity

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
	brand /model	document name /no.	minimum value	units	maximum value	units	explanation	value	units
KE1006 Data logger	Multi-Product Calibrator Fluke/5520A	In-house method: Thermocouple recorders calibration procedure (Document No.: QSZQS-42-033)	-100	°C	400	°C	K Type	0.7	°C
			-100	°C	400	°C	T Type	0.6	°C
Approval Signatory:CHIU, Yao-Ching									

Note : Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.
(Null Below)

