



財團法人全國認證基金會  
Taiwan Accreditation Foundation

**Certification Accreditation**  
(Certificate No : L2845-221230)

This is to certify that

**AUO Corporation**

**Taichung Calibration Laboratory**

No. 1, JhongKe Rd. Central Taiwan Science Park, Taichung 40763, Taiwan

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025:2017 ; CNS 17025:2018

**Accreditation Number** : 2845

**Originally Accredited** : February 20, 2014

**Effective Period** : February 20, 2023 to February 19, 2026

**Accredited Scope** : Calibration Field, see described in the Appendix



Scan to verify

*Ching-Chang Lien*

Ching-Chang Lien  
President, Taiwan Accreditation Foundation  
December 30, 2022

Accreditation Number : 2845

Laboratory Head : CHIU, Yao-Ching

## Temperature/Humidity

calibration items	working standard	calibration method	measurand level or range				measurement conditions /independent variable	smallest uncertainty	
			minimum value	units	maximum value	units		explanation	value
KE1006 Data logger	Multi-Product Calibrator Fluke/5520A	Thermocouple recorders calibration procedure (Document No.: QSZQS-42-033)	-100	°C	400	°C	K Type	0.7	°C
			-100	°C	400	°C	T Type	0.5	°C
Approval Signatory: CHIU, Yao-Ching									

Note: Smallest uncertainty represents an expanded uncertainty using a coverage factor approximately 95 % level of confidence.

(Null Below)

